


Holaday



14825 Martin Drive
Eden Prairie, Minnesota 55344-2076
Telephone: 952-934-4920
Facsimile: 952-934-3604
E-mail: info@holadayinc.com
Web page: www.holadayinc.com

CALIBRATION REPORT

ENVIRONMENT:

TEMPERATURE	23	C
-------------	----	---

LOW FREQUENCY EMF SURVEY METER(S)

					Software
MODEL:	H1-3604				
S/N:	105081				

As received, the instrument was found: _____ Within Tolerance

_____ Out of Tolerance

X New Instrument

As of this date, the above instrument(s) were calibrated to original factory specifications.

The certified accuracy of the following standards used in this calibration are traceable to the United States National Institute of Standards and Technology (NIST).

Instrument	Model	S/N	Cal Date	Due Date
Fluke True RMS Voltmeter	8060A	6516058	17 Nov 00	17 Jan 02
HP RMS Voltmeter	3400A	2415A34321	2 Oct 01	2 Oct 02
Wavetek	HD160	950713060	26 Apr 01	26 Apr 02

CR-LF-39

Calibrated: 29 Nov 01

Due: 29 Nov 01

By: PJ3/011